

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
 Data File : VV031985.D
 Acq On : 08 Sep 2023 17:00
 Operator : SY/MD
 Sample : 04277-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 09 05:36:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 09 05:33:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	261640	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	258449	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	134334	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	71592	43.722	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.440%
7) Chloroethane-d5	1.529	69	59588	41.513	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.020%
11) 1,1-Dichloroethene-d2	2.056	65	31485	41.064	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.120%
21) 2-Butanone-d5	3.793	46	131349	101.992	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.990%
24) Chloroform-d	4.253	84	141193	39.450	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.900%
26) 1,2-Dichloroethane-d4	4.947	65	95230	40.094	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.180%
32) Benzene-d6	4.963	84	293963	41.683	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.360%
36) 1,2-Dichloropropane-d6	5.992	67	98453	43.373	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.740%
41) Toluene-d8	7.246	98	265976	40.932	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.860%
43) trans-1,3-Dichloroprop...	7.558	79	47497	40.371	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.740%
47) 2-Hexanone-d5	8.030	63	86226	123.938	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	123.940%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	130706	44.560	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.120%
66) 1,2-Dichlorobenzene-d4	11.564	152	100871	41.191	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
Data File : VV031985.D
Acq On : 08 Sep 2023 17:00
Operator : SY/MD
Sample : 04277-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 09 05:36:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 09 05:33:49 2023
Response via : Initial Calibration

