

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
 Data File : VV031977.D
 Acq On : 08 Sep 2023 11:55
 Operator : SY/MD
 Sample : 04277-01 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JYCA0

Quant Time: Sep 09 05:34:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 09 05:33:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	321109	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	307760	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	160449	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	108484	53.982	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 107.960%	
7) Chloroethane-d5	1.529	69	88386	50.172	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 100.340%	
11) 1,1-Dichloroethene-d2	2.056	65	47410	50.382	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 100.760%	
21) 2-Butanone-d5	3.786	46	167741	106.128	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 106.130%	
24) Chloroform-d	4.249	84	202930	46.199	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 92.400%	
26) 1,2-Dichloroethane-d4	4.944	65	135692	46.550	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 93.100%	
32) Benzene-d6	4.960	84	429091	51.095	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 102.200%	
36) 1,2-Dichloropropane-d6	5.992	67	141942	52.513	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 105.020%	
41) Toluene-d8	7.246	98	397320	51.348	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 102.700%	
43) trans-1,3-Dichloroprop...	7.555	79	70328	50.199	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 100.400%	
47) 2-Hexanone-d5	8.027	63	118114	142.571	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 142.570%#	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	178065	50.979	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 101.960%	
66) 1,2-Dichlorobenzene-d4	11.564	152	146204	49.986	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 99.980%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
Data File : VV031977.D
Acq On : 08 Sep 2023 11:55
Operator : SY/MD
Sample : 04277-01 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JYCA0

Quant Time: Sep 09 05:34:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 09 05:33:49 2023
Response via : Initial Calibration

