

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090919\
 Data File : VV012639.D
 Acq On : 09 Sep 2019 22:41
 Operator : SY/MD
 Sample : K4780-01
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BM8

Quant Time: Sep 11 01:18:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 10 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	353775	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	357913	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	145403	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	120716	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.58	69	105457	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.20%
11) 1,1-Dichloroethene-d2	2.13	63	147604	3.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.20%
20) 2-Butanone-d5	3.95	46	370830	64.21	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.42%
24) Chloroform-d	4.40	84	229463	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	5.08	65	123366	5.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.20%
32) Benzene-d6	5.10	84	440662	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.00%
36) 1,2-Dichloropropane-d6	6.12	67	148406	5.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.40%
41) Toluene-d8	7.36	98	350924	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
43) trans-1,3-Dichloropropene-	7.66	79	46593	4.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.80%
46) 2-Hexanone-d5	8.14	63	194093	54.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.82%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	112130	5.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	146342	5.88	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090919\
Data File : VV012639.D
Acq On : 09 Sep 2019 22:41
Operator : SY/MD
Sample : K4780-01
Misc : 25mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BM8

Quant Time: Sep 11 01:18:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 10 09:27:53 2019
Response via : Initial Calibration

