

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090919\
 Data File : VV012643.D
 Acq On : 10 Sep 2019 00:28
 Operator : SY/MD
 Sample : K4703-16
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFD99

Quant Time: Sep 11 01:23:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 10 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	403900	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	409908	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	159678	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	126376	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.58	69	109495	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.13	63	151186	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	3.93	46	252417	38.28	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.56%
24) Chloroform-d	4.40	84	203243	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.60%
26) 1,2-Dichloroethane-d4	5.08	65	121958	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.10	84	435311	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.12	67	148400	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	7.36	98	356347	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.20%
43) trans-1,3-Dichloropropene-	7.66	79	45915	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
46) 2-Hexanone-d5	8.13	63	200039	49.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.84%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	107720	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	142022	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds

					Ovalue
2) Dichlorodifluoromethane	1.14	85	109406	3.212 ug/L	99
13) Acetone	2.19	43	11257	2.913 ug/L	94
17) Methyl tert-butyl Ether	2.80	73	25917	0.440 ug/L	96
22) cis-1,2-Dichloroethene	3.95	96	6430	0.210 ug/L	87
25) Chloroform	4.42	83	49033	0.838 ug/L	95
31) Carbon tetrachloride	4.87	117	9081	0.229 ug/L	98
34) Trichloroethene	5.96	95	256588	7.751 ug/L	99
47) Tetrachloroethene	8.02	164	2617293	100.716 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090919\
Data File : VV012643.D
Acq On : 10 Sep 2019 00:28
Operator : SY/MD
Sample : K4703-16
Misc : 25mL/MSVOA V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFD99

Quant Time: Sep 11 01:23:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 10 09:27:53 2019
Response via : Initial Calibration

