

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090920\
 Data File : VV018167.D
 Acq On : 09 Sep 2020 14:48
 Operator : SY/MD
 Sample : VV0909WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK63

Quant Time: Sep 10 03:44:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 10 03:30:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	326975	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	319362	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	157751	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	120400	45.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.36%
7) Chloroethane-d5	1.58	69	99952	45.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.68%
11) 1,1-Dichloroethene-d2	2.12	63	172141	35.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.44%
21) 2-Butanone-d5	3.91	46	145951	89.48	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.48%
24) Chloroform-d	4.37	84	202816	43.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.74%
26) 1,2-Dichloroethane-d4	5.05	65	135705	45.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.58%
32) Benzene-d6	5.07	84	423043	46.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.54%
36) 1,2-Dichloropropane-d6	6.09	67	134973	46.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.92%
41) Toluene-d8	7.33	98	377947	45.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.72%
43) trans-1,3-Dichloropropene-	7.64	79	65632	44.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.92%
47) 2-Hexanone-d5	8.11	63	90482	82.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.23%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	156756	42.99	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.98%
64) 1,2-Dichlorobenzene-d4	11.65	152	157187	48.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.04%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090920\
Data File : VV018167.D
Acq On : 09 Sep 2020 14:48
Operator : SY/MD
Sample : VV0909WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK63

Quant Time: Sep 10 03:44:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 10 03:30:11 2020
Response via : Initial Calibration

