

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090923\
 Data File : VV032000.D
 Acq On : 09 Sep 2023 15:04
 Operator : SY/MD
 Sample : 04267-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH4F0

Quant Time: Sep 09 21:55:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR082323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 09 21:51:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	120266	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	118218	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	62028	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	22355	5.504	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.000%
7) Chloroethane-d5	1.532	69	26987	5.800	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.000%
11) 1,1-Dichloroethene-d2	2.060	65	12944	5.415	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	108.200%
20) 2-Butanone-d5	3.806	46	107193	54.515	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.040%
24) Chloroform-d	4.256	84	64755	4.710	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.200%
26) 1,2-Dichloroethane-d4	4.950	65	34019	4.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	4.966	84	131956	4.905	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	5.995	67	43985	4.637	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.249	98	114764	4.598	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%
43) trans-1,3-Dichloroprop...	7.564	79	12365	3.575	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.400%
46) 2-Hexanone-d5	8.030	63	84276	45.696	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.400%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	31534	4.633	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	11.564	152	42531	4.234	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090923\
Data File : VV032000.D
Acq On : 09 Sep 2023 15:04
Operator : SY/MD
Sample : 04267-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH4F0

Quant Time: Sep 09 21:55:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR082323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 09 21:51:31 2023
Response via : Initial Calibration

