

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091019\
 Data File : VV012661.D
 Acq On : 10 Sep 2019 17:22
 Operator : SY/MD
 Sample : K4724-07DL 50X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFDA8DL

Quant Time: Sep 11 06:51:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 11 05:21:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	387851	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	386074	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	150647	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	111616	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.58	69	101083	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	2.13	63	135025	3.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.60%
20) 2-Butanone-d5	3.95	46	359507	56.78	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.56%
24) Chloroform-d	4.40	84	222693	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.08	65	118375	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
32) Benzene-d6	5.10	84	420369	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.12	67	144186	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.36	98	328204	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.40%
43) trans-1,3-Dichloropropene-	7.66	79	43963	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
46) 2-Hexanone-d5	8.14	63	178281	46.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	104104	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	137316	5.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.60%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	3.96	96	35482	1.206 ug/L	99
34) Trichloroethene	5.96	95	189876	6.090 ug/L	98
47) Tetrachloroethene	8.02	164	444801	18.173 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091019\
Data File : VV012661.D
Acq On : 10 Sep 2019 17:22
Operator : SY/MD
Sample : K4724-07DL 50X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFDA8DL

Quant Time: Sep 11 06:51:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 11 05:21:41 2019
Response via : Initial Calibration

