

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091020\
 Data File : VV018172.D
 Acq On : 10 Sep 2020 11:32
 Operator : SY/MD
 Sample : VV0910MBL01
 Misc : 5.0µ/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK64

Quant Time: Sep 11 01:39:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 11 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	308831	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	304468	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	147308	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	109549	43.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.06%
7) Chloroethane-d5	1.58	69	92931	45.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.24%
11) 1,1-Dichloroethene-d2	2.12	63	158962	34.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.86%
21) 2-Butanone-d5	3.91	46	142178	92.29	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.29%
24) Chloroform-d	4.38	84	188835	43.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.48%
26) 1,2-Dichloroethane-d4	5.06	65	129395	46.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.46%
32) Benzene-d6	5.08	84	398850	45.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.52%
36) 1,2-Dichloropropane-d6	6.09	67	127324	45.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.94%
41) Toluene-d8	7.34	98	354644	44.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.30%
43) trans-1,3-Dichloropropene-	7.64	79	60863	43.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.50%
47) 2-Hexanone-d5	8.11	63	88559	84.42	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.42%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	152075	43.75	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.50%
64) 1,2-Dichlorobenzene-d4	11.65	152	146314	48.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.72%

Target Compounds					Ovalue
25) Chloroform	4.41	83	9171	2.283 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091020\
Data File : VV018172.D
Acq On : 10 Sep 2020 11:32
Operator : SY/MD
Sample : VV0910MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK64

Quant Time: Sep 11 01:39:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 11 01:30:09 2020
Response via : Initial Calibration

