

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091020\
 Data File : VV018173.D
 Acq On : 10 Sep 2020 11:55
 Operator : SY/MD
 Sample : VV0910WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK65

Quant Time: Sep 11 01:42:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 11 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	295873	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	293142	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	142340	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	104876	43.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.98%
7) Chloroethane-d5	1.58	69	88754	44.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.96%
11) 1,1-Dichloroethene-d2	2.12	63	153475	34.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.40%
21) 2-Butanone-d5	3.91	46	141881	96.13	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.13%
24) Chloroform-d	4.38	84	182597	43.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.30%
26) 1,2-Dichloroethane-d4	5.06	65	126985	47.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.70%
32) Benzene-d6	5.08	84	383071	45.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.30%
36) 1,2-Dichloropropane-d6	6.09	67	124041	46.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.04%
41) Toluene-d8	7.34	98	341505	44.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.30%
43) trans-1,3-Dichloropropene-	7.64	79	59495	43.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.82%
47) 2-Hexanone-d5	8.11	63	90184	89.29	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.29%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	147876	44.19	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.38%
64) 1,2-Dichlorobenzene-d4	11.65	152	145975	49.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.88%

Target Compounds					Ovalue
25) Chloroform	4.41	83	8553	2.223 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091020\
Data File : VV018173.D
Acq On : 10 Sep 2020 11:55
Operator : SY/MD
Sample : VV0910WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK65

Quant Time: Sep 11 01:42:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 11 01:30:09 2020
Response via : Initial Calibration

