

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091020\
 Data File : VV018181.D
 Acq On : 10 Sep 2020 15:26
 Operator : SY/MD
 Sample : L3957-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWZ6

Quant Time: Sep 11 03:11:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 11 01:30:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	290037	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	289410	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	142083	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105606	44.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.36%
7) Chloroethane-d5	1.58	69	90902	47.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.12	63	154427	35.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.24%
21) 2-Butanone-d5	3.91	46	141407	97.74	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.74%
24) Chloroform-d	4.38	84	194548	47.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.88%
26) 1,2-Dichloroethane-d4	5.06	65	132243	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.60%
32) Benzene-d6	5.07	84	393660	47.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.02%
36) 1,2-Dichloropropane-d6	6.09	67	128516	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.64%
41) Toluene-d8	7.34	98	354959	47.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.02%
43) trans-1,3-Dichloropropene-	7.64	79	60953	45.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.12%
47) 2-Hexanone-d5	8.11	63	89510	89.77	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.77%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	156656	47.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.82%
64) 1,2-Dichlorobenzene-d4	11.65	152	153531	52.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.24%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091020\
Data File : VV018181.D
Acq On : 10 Sep 2020 15:26
Operator : SY/MD
Sample : L3957-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWZ6

Quant Time: Sep 11 03:11:30 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 11 01:30:09 2020
Response via : Initial Calibration

