

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM091021.M
 Title : TRACE VOA SOM01.0
 Last Update : Sat Sep 11 01:06:11 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV022134.D 0.05=VV022132.D 0.1 =VV022133.D 1.0 =VV022135.D 2 =VV022136.D

Compound		0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.689	0.869	0.640	0.708	0.693	0.720	12.15
3) T	Vinyl chloride	0.526	0.449	0.450	0.538	0.530	0.499	9.09
4) S	1,2-Dichloroetha...	0.391	0.450	0.366	0.397	0.405	0.402	7.61
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.459	0.399	0.377	0.453	0.452	0.428	8.79
7) S	1,2-Dichloroprop...	0.555	0.542	0.479	0.556	0.531	0.533	5.93
8) S	Toluene-d8	1.535	1.907	1.418	1.536	1.526	1.584	11.80
9) T	1,2-Dibromoethane	0.291	0.246	0.240	0.298	0.296	0.274	10.47
10) S	1,1,2,2-Tetrachl...	0.329	0.328	0.316	0.336	0.332	0.328	2.26
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.514	0.453	0.396	0.527	0.520	0.482	11.72
13) T	1,2-Dibromo-3-ch...	0.068	0.074	0.060	0.070	0.069	0.068	7.12

(#) = Out of Range