

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091021\
 Data File : VV022134.D
 Acq On : 10 Sep 2021 15:13
 Operator : SY/MD
 Sample : VSTD0.520
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5220

Quant Time: Sep 11 01:02:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 11 01:00:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.638	114	11263	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	10252	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	5127	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.302	65	7759	0.673	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	4409	0.569	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	5689	0.589	ug/L	0.00
8) Toluene-d8	7.316	98	15732	0.741	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.215	84	3373	0.551	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.302	62	5925	0.354	ug/L	100
6) Trichloroethene	5.925	95	4709	0.485	ug/L	100
9) 1,2-Dibromoethane	8.368	107	2988	0.533	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	2633	0.484	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.426	75	351	0.452	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091021\
Data File : VV022134.D
Acq On : 10 Sep 2021 15:13
Operator : SY/MD
Sample : VSTD0.520
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5220

Quant Time: Sep 11 01:02:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 11 01:00:11 2021
Response via : Initial Calibration

