

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091021\
 Data File : VV022136.D
 Acq On : 10 Sep 2021 16:01
 Operator : SY/MD
 Sample : VSTD00222
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002222

Quant Time: Sep 11 01:02:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091021.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 11 01:00:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	9479	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	9076	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	4537	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	26273	2.709	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.043	65	15341	2.351	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	19281	2.256	ug/L	0.00
8) Toluene-d8	7.317	98	55384	2.948	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	12048	2.223	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	20112	1.428	ug/L	95
6) Trichloroethene	5.925	95	16415	1.911	ug/L	98
9) 1,2-Dibromoethane	8.353	107	10747	2.167	ug/L #	94
12) 1,2,3-Trichloropropane	10.267	75	9443	1.963	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.427	75	1259	1.833	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091021\
Data File : VV022136.D
Acq On : 10 Sep 2021 16:01
Operator : SY/MD
Sample : VSTD00222
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002222

Quant Time: Sep 11 01:02:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091021.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 11 01:00:11 2021
Response via : Initial Calibration

