

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091119\
 Data File : VV012670.D
 Acq On : 11 Sep 2019 13:51
 Operator : SY/MD
 Sample : VV0911WBL01
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK53

Quant Time: Sep 12 07:24:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 12 07:16:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	422076	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	423585	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	169576	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	141833	4.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.00%
7) Chloroethane-d5	1.58	69	118217	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.13	63	177166	3.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.60%
20) 2-Butanone-d5	3.96	46	299911	43.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.06%
24) Chloroform-d	4.40	84	236788	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	5.08	65	118383	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
32) Benzene-d6	5.09	84	477009	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.12	67	148287	4.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.20%
41) Toluene-d8	7.36	98	381892	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.66	79	45547	4.12	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.40%
46) 2-Hexanone-d5	8.14	63	135711	32.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	64.88%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	89713	3.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	138282	4.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091119\
Data File : VV012670.D
Acq On : 11 Sep 2019 13:51
Operator : SY/MD
Sample : VV0911WBL01
Misc : 25mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK53

Quant Time: Sep 12 07:24:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 12 07:16:59 2019
Response via : Initial Calibration

