

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091119\
 Data File : VV012675.D
 Acq On : 11 Sep 2019 16:12
 Operator : SY/MD
 Sample : K4809-04DL 5X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFDC1DL

Quant Time: Sep 12 11:46:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 12 07:16:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	385538	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	387121	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	170051	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	150796	5.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.40%
7) Chloroethane-d5	1.55	69	60985	2.82	ug/L	-0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	56.40%#
11) 1,1-Dichloroethene-d2	2.12	63	191788	4.54	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.80%
20) 2-Butanone-d5	3.94	46	255496	40.60	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.20%
24) Chloroform-d	4.39	84	232750	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.08	65	131745	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.80%
32) Benzene-d6	5.09	84	523959	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.20%
36) 1,2-Dichloropropane-d6	6.11	67	164488	5.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.00%
41) Toluene-d8	7.36	98	427422	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
43) trans-1,3-Dichloropropene-	7.66	79	49325	4.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.60%
46) 2-Hexanone-d5	8.13	63	185864	48.62	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.24%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	108495	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	171971	5.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.20%

Target Compounds

					Ovalue
13) Acetone	2.19	43	11333	3.072 ug/L	97
15) Methyl Acetate	2.45	43	16670	1.680 ug/L	98
22) cis-1,2-Dichloroethene	3.95	96	188851	6.456 ug/L	100
34) Trichloroethene	5.96	95	47229	1.511 ug/L	98
47) Tetrachloroethene	8.02	164	278294	11.339 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091119\
Data File : VV012675.D
Acq On : 11 Sep 2019 16:12
Operator : SY/MD
Sample : K4809-04DL 5X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFDC1DL

Quant Time: Sep 12 11:46:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 12 07:16:59 2019
Response via : Initial Calibration

