

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091123\
 Data File : VW032025.D
 Acq On : 11 Sep 2023 14:42
 Operator : SY/MD
 Sample : 04286-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 12 06:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR082323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 09 21:51:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	128363	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	127335	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	66361	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	21858	5.042	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.800%
7) Chloroethane-d5	1.532	69	27126	5.462	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.200%
11) 1,1-Dichloroethene-d2	2.060	65	12817	5.023	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.400%
20) 2-Butanone-d5	3.812	46	109735	52.288	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.580%
24) Chloroform-d	4.256	84	66635	4.541	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	4.950	65	35593	4.708	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	4.963	84	136214	4.701	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	5.995	67	47232	4.623	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.249	98	115539	4.297	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.000%
43) trans-1,3-Dichloroprop...	7.564	79	12509	3.357	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.200%
46) 2-Hexanone-d5	8.030	63	84868	42.722	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.440%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	32928	4.492	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	43729	4.069	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091123\
Data File : VV032025.D
Acq On : 11 Sep 2023 14:42
Operator : SY/MD
Sample : 04286-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 12 06:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR082323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 09 21:51:31 2023
Response via : Initial Calibration

