

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091222\
 Data File : VV027805.D
 Acq On : 12 Sep 2022 12:39
 Operator : SY/MD
 Sample : N3670-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT2-2022-03

Quant Time: Sep 13 03:23:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 10 02:33:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	3034	0.500	ug/L	0.02
5) Chlorobenzene-d5	8.863	117	2843	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	1995	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	926	0.492	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.000%
4) 1,2-Dichloroethane-d4	5.042	65	817	0.577	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	116.000%
7) 1,2-Dichloropropane-d6	6.068	67	538	0.568	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	114.000%
8) Toluene-d8	7.317	98	2965	0.566	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	114.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	551	0.679	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	136.000%#
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	43	0.024	ug/L #	1
6) Trichloroethene	5.925	95	37	0.021	ug/L #	67
9) 1,2-Dibromoethane	8.368	107	23	0.022	ug/L #	69
12) 1,2,3-Trichloropropane	10.267	75	23	0.035	ug/L #	55
13) 1,2-Dibromo-3-chloropr...	12.426	75	8	0.061	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091222\
Data File : VV027805.D
Acq On : 12 Sep 2022 12:39
Operator : SY/MD
Sample : N3670-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT2-2022-03

Quant Time: Sep 13 03:23:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 10 02:33:14 2022
Response via : Initial Calibration

