

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091222\
 Data File : VV027803.D
 Acq On : 12 Sep 2022 10:37
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5285

Quant Time: Sep 13 03:21:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 10 02:33:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.615	114	4437	0.500 ug/L	0.00
5) Chlorobenzene-d5	8.848	117	3958	0.500 ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	2743	0.500 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.296	65	1461	0.531 ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.000%		
4) 1,2-Dichloroethane-d4	5.018	65	1124	0.543 ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%		
7) 1,2-Dichloropropane-d6	6.068	67	738	0.560 ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 112.000%		
8) Toluene-d8	7.301	98	4069	0.558 ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery = 112.000%		
10) 1,1,2,2-Tetrachloroeth...	10.198	84	639	0.566 ug/L	-0.02
Spiked Amount 0.500	Range 65 - 120		Recovery = 114.000%		
Target Compounds					
					Qvalue
3) Vinyl chloride	1.303	62	1338	0.518 ug/L	97
6) Trichloroethene	5.901	95	1323	0.536 ug/L	99
9) 1,2-Dibromoethane	8.353	107	759	0.527 ug/L	97
12) 1,2,3-Trichloropropane	10.267	75	457	0.508 ug/L	92
13) 1,2-Dibromo-3-chloropr...	12.409	75	110	0.607 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091222\
Data File : VV027803.D
Acq On : 12 Sep 2022 10:37
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5285

Quant Time: Sep 13 03:21:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 10 02:33:14 2022
Response via : Initial Calibration

