

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091223\
 Data File : VW032047.D
 Acq On : 12 Sep 2023 14:41
 Operator : SY/MD
 Sample : 04385-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EZYP3

Quant Time: Sep 13 06:56:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 09 05:33:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	243702	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	238421	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	116940	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	68596	44.976	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.960%
7) Chloroethane-d5	1.529	69	56852	42.522	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.040%
11) 1,1-Dichloroethene-d2	2.060	65	28952	40.540	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.080%
21) 2-Butanone-d5	3.789	46	122825	102.393	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.390%
24) Chloroform-d	4.256	84	127094	38.125	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.240%
26) 1,2-Dichloroethane-d4	4.947	65	86356	39.034	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.060%
32) Benzene-d6	4.963	84	268150	41.217	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.440%
36) 1,2-Dichloropropane-d6	5.992	67	91781	43.830	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.660%
41) Toluene-d8	7.246	98	237059	39.546	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.100%#
43) trans-1,3-Dichloroprop...	7.558	79	39876	36.741	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.480%
47) 2-Hexanone-d5	8.030	63	72630	113.165	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.170%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	114506	42.316	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.640%
66) 1,2-Dichlorobenzene-d4	11.564	152	89600	42.031	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.060%
Target Compounds					Qvalue	
13) Acetone	2.114	43	11543	10.811	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091223\
Data File : VV032047.D
Acq On : 12 Sep 2023 14:41
Operator : SY/MD
Sample : 04385-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EZYP3

Quant Time: Sep 13 06:56:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 09 05:33:49 2023
Response via : Initial Calibration

