

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091225\
 Data File : VV039053.D
 Acq On : 12 Sep 2025 11:50
 Operator : SY/MD
 Sample : VX0912WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK271

Quant Time: Sep 13 00:27:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM091125WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 12 01:49:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	805389	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.773	117	822870	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.175	152	324716	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.291	65	341064	46.099	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.200%
7) Chloroethane-d5	1.545	69	287940	49.202	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.400%
11) 1,1-Dichloroethene-d2	2.069	65	155035	48.583	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.160%
21) 2-Butanone-d5	3.786	46	309268	106.521	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.520%
24) Chloroform-d	4.253	84	600093	49.182	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.360%
26) 1,2-Dichloroethane-d4	4.941	65	363497	50.588	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.180%
32) Benzene-d6	4.960	84	1152498	48.312	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.620%
36) 1,2-Dichloropropane-d6	5.986	67	370552	47.475	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.960%
41) Toluene-d8	7.236	98	969484	43.533	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.060%
43) trans-1,3-Dichloroprop...	7.548	79	164159	44.502	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.000%
47) 2-Hexanone-d5	8.014	63	230157	103.585	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.590%
56) 1,1,2,2-Tetrachloroeth...	10.140	84	535602	48.188	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.380%
66) 1,2-Dichlorobenzene-d4	11.548	152	333624	54.705	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091225\
Data File : VV039053.D
Acq On : 12 Sep 2025 11:50
Operator : SY/MD
Sample : VX0912WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK271

Quant Time: Sep 13 00:27:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM091125WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 12 01:49:00 2025
Response via : Initial Calibration

