

Method Path : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\
 Method File : SOMVLM091319WMA.M
 Title : VOC Analysis
 Last Update : Fri Sep 13 18:43:16 2019
 Response Via : Initial Calibration

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB70

Calibration Files

5 =VV012732.D 10 =VV012731.D 50 =VV012727.D
 100 =VV012728.D 200 =VV012729.D

	Compound	5	10	50	100	200	Avg	%RSD
53) T	m,p-Xylene	0.615	0.410	0.600	0.608	0.627	0.572	15.96
54) T	o-xylene	0.621	0.420	0.607	0.615	0.639	0.580	15.54
55) T	Styrene	0.983	0.690	1.065	1.069	1.115	0.985	17.39
56) T	Isopropylbenzene	1.627	1.081	1.619	1.619	1.680	1.525	16.37
57) S	1,1,2,2-Tetrachloro	0.552	0.572	0.570	0.645	0.672	0.602	8.75
58) T	1,1,2,2-Tetrachloro	0.594	0.394	0.510	0.536	0.571	0.521	14.92
59)	1,2,3-Trichloroprop	0.607	0.410	0.485	0.487	0.500	0.498	14.22
60) I	1,4-Dichlorobenzene-d	-----ISTD-----						
61) T	Bromoform	0.745	0.506	0.570	0.586	0.612	0.604	14.61
62) T	1,3-Dichlorobenzene	1.879	1.267	1.535	1.507	1.533	1.544	14.12
63) T	1,4-Dichlorobenzene	1.982	1.336	1.550	1.512	1.541	1.584	15.06
64) S	1,2-Dichlorobenzene	1.132	1.181	1.079	1.152	1.162	1.141	3.40
65) T	1,2-Dichlorobenzene	2.030	1.359	1.588	1.563	1.580	1.624	15.14
66) T	1,2-Dibromo-3-chlor	0.255	0.164	0.196	0.222	0.246	0.217	17.32
67)	1,3,5-Trichlorobenz	1.352	0.908	1.124	1.122	1.199	1.141	14.05
68) T	1,2,4-trichlorobenz	1.067	0.711	0.939	0.997	1.105	0.964	16.11
69)	Naphthalene	2.591	1.815	2.905	3.204	3.419	2.787	22.47
70) T	1,2,3-Trichlorobenz	1.159	0.795	1.073	1.070	1.135	1.046	13.92

(#) = Out of Range