

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091319\
 Data File : VV012776.D
 Acq On : 14 Sep 2019 19:22
 Operator : SY/MD
 Sample : K4867-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMG8

Quant Time: Sep 16 05:55:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM091319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 16 04:57:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	689032	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	640355	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	221073	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	158180	40.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.02%
7) Chloroethane-d5	1.58	69	164714	47.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.26%
11) 1,1-Dichloroethene-d2	2.13	63	250316	32.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.90%
21) 2-Butanone-d5	3.92	46	450237	111.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.18%
24) Chloroform-d	4.40	84	482688	49.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.78%
26) 1,2-Dichloroethane-d4	5.08	65	316853	53.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.96%
32) Benzene-d6	5.10	84	878851	51.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.86%
36) 1,2-Dichloropropane-d6	6.12	67	323810	53.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.42%
41) Toluene-d8	7.36	98	722681	46.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.46%
43) trans-1,3-Dichloropropene-	7.66	79	121543	48.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.76%
47) 2-Hexanone-d5	8.13	63	230393	99.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.59%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	387864	50.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.56%
64) 1,2-Dichlorobenzene-d4	11.67	152	289119	57.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091319\
Data File : VV012776.D
Acq On : 14 Sep 2019 19:22
Operator : SY/MD
Sample : K4867-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0MG8

Quant Time: Sep 16 05:55:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM091319WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 16 04:57:02 2019
Response via : Initial Calibration

