

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091319\
 Data File : VV012763.D
 Acq On : 14 Sep 2019 12:52
 Operator : SY/MD
 Sample : K4707-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 16 04:54:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM091319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 16 04:18:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	728998	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	693360	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	265395	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	176178	42.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.28%
7) Chloroethane-d5	1.58	69	171849	46.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.96%
11) 1,1-Dichloroethene-d2	2.13	63	274984	34.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.44%
21) 2-Butanone-d5	3.92	46	438979	102.46	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.46%
24) Chloroform-d	4.40	84	484246	47.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	5.08	65	315167	50.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.50%
32) Benzene-d6	5.10	84	904115	48.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.72%
36) 1,2-Dichloropropane-d6	6.12	67	329191	50.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.84%
41) Toluene-d8	7.36	98	782217	46.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.42%
43) trans-1,3-Dichloropropene-	7.66	79	124041	45.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.20%
47) 2-Hexanone-d5	8.13	63	225490	90.02	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.02%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	373643	44.73	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.46%
64) 1,2-Dichlorobenzene-d4	11.67	152	336648	55.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091319\
Data File : VV012763.D
Acq On : 14 Sep 2019 12:52
Operator : SY/MD
Sample : K4707-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 16 04:54:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM091319WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 16 04:18:41 2019
Response via : Initial Calibration

