

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091319\
 Data File : VV012770.D
 Acq On : 14 Sep 2019 16:56
 Operator : SY/MD
 Sample : K4868-06
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COML1

Quant Time: Sep 16 05:38:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM091319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 16 04:57:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	698604	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	651226	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	232003	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	164652	41.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.18%
7) Chloroethane-d5	1.58	69	164395	46.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.80%
11) 1,1-Dichloroethene-d2	2.13	63	259733	33.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.46%
21) 2-Butanone-d5	3.92	46	437525	106.56	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.56%
24) Chloroform-d	4.40	84	473699	48.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.58%
26) 1,2-Dichloroethane-d4	5.08	65	305972	51.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.84%
32) Benzene-d6	5.10	84	880665	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.36%
36) 1,2-Dichloropropane-d6	6.12	67	319697	52.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.28%
41) Toluene-d8	7.36	98	743312	46.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.52%
43) trans-1,3-Dichloropropene-	7.66	79	120442	47.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.28%
47) 2-Hexanone-d5	8.13	63	223112	94.83	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.83%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	374109	47.69	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.38%
64) 1,2-Dichlorobenzene-d4	11.67	152	296724	56.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091319\
Data File : VV012770.D
Acq On : 14 Sep 2019 16:56
Operator : SY/MD
Sample : K4868-06
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COML1

Quant Time: Sep 16 05:38:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM091319WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 16 04:57:02 2019
Response via : Initial Calibration

