

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091322\
 Data File : VV027815.D
 Acq On : 13 Sep 2022 13:33
 Operator : SY/MD
 Sample : N4519-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT4-2022-08

Quant Time: Sep 14 02:31:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 10 02:33:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	3924	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.848	117	3484	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	2422	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.296	65	1241	0.510	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	5.018	65	989	0.540	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 108.000%	
7) 1,2-Dichloropropane-d6	6.068	67	624	0.538	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.301	98	3462	0.539	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 108.000%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	539	0.542	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.296	62	46	0.020	ug/L #	1
6) Trichloroethene	5.901	95	39	0.018	ug/L #	69
9) 1,2-Dibromoethane	8.353	107	18	0.014	ug/L #	67
12) 1,2,3-Trichloropropane	10.267	75	13	0.016	ug/L #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091322\
Data File : VV027815.D
Acq On : 13 Sep 2022 13:33
Operator : SY/MD
Sample : N4519-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT4-2022-08

Quant Time: Sep 14 02:31:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 10 02:33:14 2022
Response via : Initial Calibration

