

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091322\
 Data File : VV027816.D
 Acq On : 13 Sep 2022 14:08
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5288

Quant Time: Sep 14 02:31:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 10 02:33:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	4073	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	3761	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	2599	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.296	65	1302	0.516	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	104.000%
4) 1,2-Dichloroethane-d4	5.018	65	1001	0.527	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	6.068	67	656	0.524	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.000%
8) Toluene-d8	7.317	98	3473	0.501	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	549	0.512	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.296	62	1238	0.522	ug/L	98
6) Trichloroethene	5.901	95	1191	0.508	ug/L	91
9) 1,2-Dibromoethane	8.353	107	696	0.509	ug/L	99
12) 1,2,3-Trichloropropane	10.267	75	431	0.505	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.409	75	94	0.547	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091322\
Data File : VV027816.D
Acq On : 13 Sep 2022 14:08
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5288

Quant Time: Sep 14 02:31:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 10 02:33:14 2022
Response via : Initial Calibration

