

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091418\
 Data File : VV007552.D
 Acq On : 15 Sep 2018 04:31
 Operator : SY/MD
 Sample : J4872-14
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KR6

Quant Time: Sep 15 06:46:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 15 02:08:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	118885	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	103925	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	46045	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	24217	3.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.00%
7) Chloroethane-d5	1.58	69	24802	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.13	63	44329	3.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.60%
20) 2-Butanone-d5	3.97	46	38380	44.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.62%
24) Chloroform-d	4.41	84	51553	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
26) 1,2-Dichloroethane-d4	5.09	65	27163	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
32) Benzene-d6	5.10	84	104475	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.13	67	31115	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.36	98	102742	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	7.67	79	10658	3.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.60%
46) 2-Hexanone-d5	8.15	63	30291	58.35	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.70%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20603	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	39215	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	5826	0.885 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091418\
Data File : VV007552.D
Acq On : 15 Sep 2018 04:31
Operator : SY/MD
Sample : J4872-14
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KR6

Quant Time: Sep 15 06:46:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 15 02:08:52 2018
Response via : Initial Calibration

