

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV091418\
 Data File : VV007554.D
 Acq On : 15 Sep 2018 05:27
 Operator : SY/MD
 Sample : J4872-16
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KS2

Quant Time: Sep 19 04:48:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 15 02:08:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	119259	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	101977	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	46027	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25535	4.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.00%
7) Chloroethane-d5	1.58	69	27263	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.13	63	45268	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.97	46	56556	65.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	130.18%#
24) Chloroform-d	4.41	84	55766	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.09	65	27946	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
32) Benzene-d6	5.10	84	108804	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	6.13	67	31862	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	7.36	98	107104	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloropropene-	7.67	79	10551	3.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.20%
46) 2-Hexanone-d5	8.14	63	34654	68.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	136.06%#
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20140	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	39221	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV091418\
Data File : VV007554.D
Acq On : 15 Sep 2018 05:27
Operator : SY/MD
Sample : J4872-16
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KS2

Quant Time: Sep 19 04:48:34 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 15 02:08:52 2018
Response via : Initial Calibration

