

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091418\
 Data File : VV007558.D
 Acq On : 15 Sep 2018 07:19
 Operator : SY/MD
 Sample : J4872-20
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KT3

Quant Time: Sep 17 06:58:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 04:48:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	106706	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	88178	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	38442	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	21623	3.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.60%
7) Chloroethane-d5	1.58	69	22717	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.13	63	39172	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	3.97	46	43574	56.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.10%
24) Chloroform-d	4.41	84	49874	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.09	65	26661	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
32) Benzene-d6	5.10	84	96692	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.13	67	30017	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.36	98	90513	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
43) trans-1,3-Dichloropropene-	7.67	79	9852	3.57	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.40%
46) 2-Hexanone-d5	8.15	63	26826	60.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.80%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	19257	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	32860	4.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091418\
Data File : VV007558.D
Acq On : 15 Sep 2018 07:19
Operator : SY/MD
Sample : J4872-20
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KT3

Quant Time: Sep 17 06:58:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 04:48:21 2018
Response via : Initial Calibration

