

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091418\
 Data File : VV007559.D
 Acq On : 15 Sep 2018 07:48
 Operator : SY/MD
 Sample : J4872-21
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KT6

Quant Time: Sep 17 07:08:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 04:48:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	113955	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	98978	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	43917	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	24040	4.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.80%
7) Chloroethane-d5	1.58	69	23904	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.80%
11) 1,1-Dichloroethene-d2	2.13	63	43227	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.98	46	39631	47.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.46%
24) Chloroform-d	4.41	84	53454	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.09	65	26558	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
32) Benzene-d6	5.10	84	106581	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.12	67	30692	4.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.80%
41) Toluene-d8	7.36	98	101301	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
43) trans-1,3-Dichloropropene-	7.67	79	10314	3.33	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.60%
46) 2-Hexanone-d5	8.15	63	28473	57.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.18%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	19573	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	37869	4.71	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.20%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	4907	0.783 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091418\
Data File : VV007559.D
Acq On : 15 Sep 2018 07:48
Operator : SY/MD
Sample : J4872-21
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KT6

Quant Time: Sep 17 07:08:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 04:48:21 2018
Response via : Initial Calibration

