

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV091418\
 Data File : VV007555.D
 Acq On : 15 Sep 2018 05:55
 Operator : SY/MD
 Sample : J4872-17
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KS5

Quant Time: Sep 19 04:48:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 15 02:08:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	113215	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	94498	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	41740	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	23046	3.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.00%
7) Chloroethane-d5	1.58	69	28497	5.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.00%
11) 1,1-Dichloroethene-d2	2.13	63	41034	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	3.97	46	55767	67.61	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	135.22%#
24) Chloroform-d	4.41	84	52293	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.09	65	27589	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
32) Benzene-d6	5.10	84	103044	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.12	67	31080	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.36	98	97532	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	7.67	79	10172	3.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.80%
46) 2-Hexanone-d5	8.15	63	32522	68.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	137.78%#
57) 1,1,2,2-Tetrachloroethane-	10.27	84	19806	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	35050	4.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV091418\
Data File : VV007555.D
Acq On : 15 Sep 2018 05:55
Operator : SY/MD
Sample : J4872-17
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KS5

Quant Time: Sep 19 04:48:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 15 02:08:52 2018
Response via : Initial Calibration

