

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091420\
 Data File : VV018201.D
 Acq On : 14 Sep 2020 14:07
 Operator : SY/MD
 Sample : VV0914WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK68

Quant Time: Sep 15 05:48:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 15 05:21:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	289615	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	287360	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	141803	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	101520	43.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.02%
7) Chloroethane-d5	1.58	69	92094	47.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.36%
11) 1,1-Dichloroethene-d2	2.12	63	152782	35.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.58%
21) 2-Butanone-d5	3.91	46	146605	101.48	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.48%
24) Chloroform-d	4.37	84	201671	49.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.50%
26) 1,2-Dichloroethane-d4	5.05	65	133162	50.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.46%
32) Benzene-d6	5.07	84	401719	48.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.66%
36) 1,2-Dichloropropane-d6	6.09	67	130280	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.68%
41) Toluene-d8	7.33	98	356537	47.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.12%
43) trans-1,3-Dichloropropene-	7.64	79	64789	48.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.56%
47) 2-Hexanone-d5	8.10	63	111363	112.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.48%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	175177	53.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	155002	53.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091420\
Data File : VV018201.D
Acq On : 14 Sep 2020 14:07
Operator : SY/MD
Sample : VV0914WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK68

Quant Time: Sep 15 05:48:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 15 05:21:42 2020
Response via : Initial Calibration

