

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091420\
 Data File : VV018203.D
 Acq On : 14 Sep 2020 15:05
 Operator : SY/MD
 Sample : L4012-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFX06

Quant Time: Sep 15 05:58:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 15 05:21:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	293935	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	291601	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	141708	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94920	39.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.24%
7) Chloroethane-d5	1.58	69	87712	44.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.50%
11) 1,1-Dichloroethene-d2	2.12	63	144578	32.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.80%
21) 2-Butanone-d5	3.91	46	146376	99.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.83%
24) Chloroform-d	4.38	84	199399	47.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.96%
26) 1,2-Dichloroethane-d4	5.06	65	134235	50.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.78%
32) Benzene-d6	5.08	84	394285	47.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.46%
36) 1,2-Dichloropropane-d6	6.09	67	127881	48.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.42%
41) Toluene-d8	7.34	98	346877	45.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	7.64	79	64191	47.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.24%
47) 2-Hexanone-d5	8.11	63	111177	110.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.66%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	173481	52.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	150643	51.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.52%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091420\
Data File : VV018203.D
Acq On : 14 Sep 2020 15:05
Operator : SY/MD
Sample : L4012-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFX06

Quant Time: Sep 15 05:58:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 15 05:21:42 2020
Response via : Initial Calibration

