

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091420\
 Data File : VV018202.D
 Acq On : 14 Sep 2020 14:43
 Operator : SY/MD
 Sample : L4012-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWZ7

Quant Time: Sep 15 05:55:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 15 05:21:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	289389	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	290261	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	142463	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	98270	41.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.34%
7) Chloroethane-d5	1.58	69	90829	47.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.14%
11) 1,1-Dichloroethene-d2	2.12	63	150938	34.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.78%
21) 2-Butanone-d5	3.91	46	149506	103.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.57%
24) Chloroform-d	4.37	84	203923	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.68%
26) 1,2-Dichloroethane-d4	5.05	65	134943	51.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.90%
32) Benzene-d6	5.07	84	400015	48.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.28%
36) 1,2-Dichloropropane-d6	6.09	67	131552	49.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.66%
41) Toluene-d8	7.33	98	355999	47.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.02%
43) trans-1,3-Dichloropropene-	7.64	79	63900	47.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.26%
47) 2-Hexanone-d5	8.11	63	113251	113.24	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.24%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	178602	53.90	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	154473	52.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091420\
Data File : VV018202.D
Acq On : 14 Sep 2020 14:43
Operator : SY/MD
Sample : L4012-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWZ7

Quant Time: Sep 15 05:55:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 15 05:21:42 2020
Response via : Initial Calibration

