

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
 Data File : VV027821.D
 Acq On : 14 Sep 2022 11:54
 Operator : SY/MD
 Sample : N4546-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG6H4

Quant Time: Sep 15 02:49:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	129737	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.863	117	111578	5.000	ug/L	0.02
58) 1,4-Dichlorobenzene-d4	11.249	152	51119	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	43995	5.144	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.800%
7) Chloroethane-d5	1.568	69	43218	5.957	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.200%
11) 1,1-Dichloroethene-d2	2.104	63	76064	4.192	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.800%
20) 2-Butanone-d5	3.963	46	115558	60.973	ug/L	0.07
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.940%
24) Chloroform-d	4.352	84	108111	6.174	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	123.400%
26) 1,2-Dichloroethane-d4	5.037	65	61729	6.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	132.400%#
32) Benzene-d6	5.040	84	215704	6.664	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	133.200%#
36) 1,2-Dichloropropane-d6	6.072	67	54846	5.787	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.800%
41) Toluene-d8	7.333	98	148690	5.463	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.200%
43) trans-1,3-Dichloroprop...	7.644	79	15938	5.013	ug/L	0.03
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.200%
46) 2-Hexanone-d5	8.114	63	59139	50.040	ug/L	0.03
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.080%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	36607	6.015	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.200%#
66) 1,2-Dichlorobenzene-d4	11.625	152	53052	6.318	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	126.400%#
Target Compounds						Qvalue
47) Tetrachloroethene	7.989	164	2977	0.402	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027821.D
Acq On : 14 Sep 2022 11:54
Operator : SY/MD
Sample : N4546-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG6H4

Quant Time: Sep 15 02:49:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

