

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091422\
 Data File : VW027839.D
 Acq On : 14 Sep 2022 19:38
 Operator : SY/MD
 Sample : N4618-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E45

Quant Time: Sep 15 02:53:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	120696	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	123333	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58839	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31509	3.960	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.200%
7) Chloroethane-d5	1.568	69	32446	4.807	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.200%
11) 1,1-Dichloroethene-d2	2.105	63	53943	3.195	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.000%
20) 2-Butanone-d5	3.895	46	40379	22.901	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	45.800%
24) Chloroform-d	4.346	84	76356	4.687	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	5.030	65	40336	4.650	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
32) Benzene-d6	5.047	84	142581	3.985	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.600%
36) 1,2-Dichloropropane-d6	6.069	67	41213	3.934	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.600%
41) Toluene-d8	7.317	98	118467	3.938	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	7.625	79	13293	3.782	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.600%
46) 2-Hexanone-d5	8.091	63	49705	38.049	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.100%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	28587	4.250	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	43712	4.522	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027839.D
Acq On : 14 Sep 2022 19:38
Operator : SY/MD
Sample : N4618-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E45

Quant Time: Sep 15 02:53:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

