

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091422\
 Data File : VW027830.D
 Acq On : 14 Sep 2022 15:57
 Operator : SY/MD
 Sample : N4618-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E15

Quant Time: Sep 15 02:51:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	123490	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	126005	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	59791	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33068	4.062	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.200%
7) Chloroethane-d5	1.568	69	32207	4.664	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.200%
11) 1,1-Dichloroethene-d2	2.105	63	55106	3.190	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.800%
20) 2-Butanone-d5	3.892	46	41645	23.085	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	46.160%
24) Chloroform-d	4.346	84	83904	5.034	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.031	65	41807	4.710	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	5.047	84	151628	4.148	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.066	67	44375	4.146	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.000%
41) Toluene-d8	7.314	98	121759	3.961	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	7.625	79	14276	3.976	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.092	63	52221	39.127	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.260%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29315	4.265	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	45593	4.642	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027830.D
Acq On : 14 Sep 2022 15:57
Operator : SY/MD
Sample : N4618-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E15

Quant Time: Sep 15 02:51:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

