

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
 Data File : VV027832.D
 Acq On : 14 Sep 2022 16:46
 Operator : SY/MD
 Sample : N4618-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E21

Quant Time: Sep 15 02:52:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	124060	5.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	122537	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59601	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	31964	3.909	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.200%
7) Chloroethane-d5	1.564	69	30820	4.443	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.800%
11) 1,1-Dichloroethene-d2	2.105	63	51524	2.969	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.400%#
20) 2-Butanone-d5	3.889	46	37424	20.650	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	41.300%
24) Chloroform-d	4.342	84	79441	4.744	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	5.031	65	39304	4.408	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
32) Benzene-d6	5.047	84	145788	4.101	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
36) 1,2-Dichloropropane-d6	6.069	67	42971	4.129	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.600%
41) Toluene-d8	7.313	98	116949	3.912	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
43) trans-1,3-Dichloroprop...	7.625	79	13365	3.828	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.600%
46) 2-Hexanone-d5	8.091	63	48920	37.691	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.380%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	27380	4.097	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	43800	4.474	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027832.D
Acq On : 14 Sep 2022 16:46
Operator : SY/MD
Sample : N4618-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E21

Quant Time: Sep 15 02:52:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

