

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091422\
 Data File : VW027836.D
 Acq On : 14 Sep 2022 18:24
 Operator : SY/MD
 Sample : N4618-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E33

Quant Time: Sep 15 02:53:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	121763	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	122146	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	57974	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33625	4.189	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%
7) Chloroethane-d5	1.568	69	32438	4.764	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.108	63	57552	3.379	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	3.889	46	55631	31.275	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	62.560%
24) Chloroform-d	4.346	84	84965	5.170	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.400%
26) 1,2-Dichloroethane-d4	5.031	65	42012	4.800	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	5.050	84	155143	4.378	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	6.066	67	43940	4.235	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.800%
41) Toluene-d8	7.313	98	126852	4.257	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	7.625	79	14228	4.088	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.800%
46) 2-Hexanone-d5	8.091	63	55578	42.958	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.920%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	30204	4.534	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	48162	5.057	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027836.D
Acq On : 14 Sep 2022 18:24
Operator : SY/MD
Sample : N4618-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E33

Quant Time: Sep 15 02:53:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

