

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
 Data File : VV027837.D
 Acq On : 14 Sep 2022 18:48
 Operator : SY/MD
 Sample : N4618-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E36

Quant Time: Sep 15 02:53:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	126223	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	122909	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58528	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31905	3.835	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.600%
7) Chloroethane-d5	1.568	69	30259	4.287	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.800%
11) 1,1-Dichloroethene-d2	2.105	63	53362	3.022	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.400%
20) 2-Butanone-d5	3.896	46	45613	24.737	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	49.480%
24) Chloroform-d	4.346	84	79414	4.661	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.034	65	39970	4.406	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
32) Benzene-d6	5.047	84	145747	4.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
36) 1,2-Dichloropropane-d6	6.069	67	40211	3.852	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	77.000%
41) Toluene-d8	7.313	98	121057	4.037	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	7.625	79	12917	3.688	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.091	63	51863	39.837	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.680%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28471	4.247	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	46151	4.800	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027837.D
Acq On : 14 Sep 2022 18:48
Operator : SY/MD
Sample : N4618-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E36

Quant Time: Sep 15 02:53:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

