

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091422\
 Data File : VW027838.D
 Acq On : 14 Sep 2022 19:13
 Operator : SY/MD
 Sample : N4618-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E39

Quant Time: Sep 15 02:53:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 15 02:47:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	123867	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	121488	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	57842	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31248	3.827	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.600%
7) Chloroethane-d5	1.568	69	31928	4.610	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.108	63	54389	3.139	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.800%
20) 2-Butanone-d5	3.899	46	43846	24.231	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	48.460%
24) Chloroform-d	4.346	84	83410	4.989	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.034	65	41295	4.638	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
32) Benzene-d6	5.050	84	151615	4.302	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
36) 1,2-Dichloropropane-d6	6.066	67	41770	4.048	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.000%
41) Toluene-d8	7.313	98	120662	4.071	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.400%
43) trans-1,3-Dichloroprop...	7.625	79	13028	3.763	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%
46) 2-Hexanone-d5	8.091	63	51472	40.000	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.000%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28560	4.310	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	46282	4.871	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091422\
Data File : VV027838.D
Acq On : 14 Sep 2022 19:13
Operator : SY/MD
Sample : N4618-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E39

Quant Time: Sep 15 02:53:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 15 02:47:38 2022
Response via : Initial Calibration

