

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091518\
 Data File : VV007575.D
 Acq On : 15 Sep 2018 19:47
 Operator : SY/MD
 Sample : J4888-03
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KW0

Quant Time: Sep 17 03:06:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 01:45:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	119430	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	106336	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	54852	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25001	4.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.20%
7) Chloroethane-d5	1.59	69	24042	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.00%
11) 1,1-Dichloroethene-d2	2.14	63	44778	3.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.00%
20) 2-Butanone-d5	3.97	46	51965	59.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.44%
24) Chloroform-d	4.41	84	58422	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.09	65	29833	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
32) Benzene-d6	5.10	84	114620	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.13	67	34541	4.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.00%
41) Toluene-d8	7.36	98	115675	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
43) trans-1,3-Dichloropropene-	7.67	79	13357	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.15	63	21083	39.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.38%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21593	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	48677	4.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.00%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	8252	1.23 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091518\
Data File : VV007575.D
Acq On : 15 Sep 2018 19:47
Operator : SY/MD
Sample : J4888-03
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KW0

Quant Time: Sep 17 03:06:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 01:45:59 2018
Response via : Initial Calibration

