

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091519\
 Data File : VV012794.D
 Acq On : 15 Sep 2019 11:42
 Operator : SY/MD
 Sample : K4834-19
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMM1

Quant Time: Sep 17 11:24:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 07:16:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	432023	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	425663	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	160856	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	184852	6.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	126.20%
7) Chloroethane-d5	1.58	69	156219	6.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	129.00%
11) 1,1-Dichloroethene-d2	2.13	63	229142	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.80%
20) 2-Butanone-d5	3.96	46	378672	53.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.40%
24) Chloroform-d	4.40	84	308362	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.80%
26) 1,2-Dichloroethane-d4	5.08	65	155856	6.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.20%
32) Benzene-d6	5.10	84	615812	6.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.00%
36) 1,2-Dichloropropane-d6	6.12	67	192720	5.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.40%
41) Toluene-d8	7.36	98	494622	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.60%
43) trans-1,3-Dichloropropene-	7.66	79	61265	5.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	110.40%
46) 2-Hexanone-d5	8.14	63	198589	47.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.48%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	110523	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	165899	6.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.60%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091519\
Data File : VV012794.D
Acq On : 15 Sep 2019 11:42
Operator : SY/MD
Sample : K4834-19
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMM1

Quant Time: Sep 17 11:24:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 07:16:11 2019
Response via : Initial Calibration

