

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091519\
 Data File : VV012805.D
 Acq On : 15 Sep 2019 19:16
 Operator : SY/MD
 Sample : K4834-15
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMJ3

Quant Time: Sep 16 08:29:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 08:14:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	377634	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	366626	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	132539	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	194002	7.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	151.40%#
7) Chloroethane-d5	1.58	69	168444	7.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	159.00%#
11) 1,1-Dichloroethene-d2	2.13	63	236739	5.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.40%
20) 2-Butanone-d5	3.99	46	435592	70.66	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	141.32%#
24) Chloroform-d	4.40	84	327167	7.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	140.60%#
26) 1,2-Dichloroethane-d4	5.08	65	177051	7.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	159.00%#
32) Benzene-d6	5.09	84	661948	7.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	151.00%#
36) 1,2-Dichloropropane-d6	6.12	67	208569	7.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	150.00%#
41) Toluene-d8	7.36	98	519861	6.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	132.40%#
43) trans-1,3-Dichloropropene-	7.66	79	65551	6.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	137.00%#
46) 2-Hexanone-d5	8.14	63	233087	64.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	128.76%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	122222	6.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	122.40%#
64) 1,2-Dichlorobenzene-d4	11.67	152	177380	7.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	156.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091519\
Data File : VV012805.D
Acq On : 15 Sep 2019 19:16
Operator : SY/MD
Sample : K4834-15
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMJ3

Quant Time: Sep 16 08:29:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 08:14:10 2019
Response via : Initial Calibration

