

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091519\
 Data File : VV012808.D
 Acq On : 15 Sep 2019 20:35
 Operator : SY/MD
 Sample : K4865-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMJ8

Quant Time: Sep 16 08:37:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 08:14:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	401651	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	392163	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	136952	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	203588	7.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	149.40%#
7) Chloroethane-d5	1.58	69	169783	7.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	150.60%#
11) 1,1-Dichloroethene-d2	2.13	63	253333	5.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	115.20%
20) 2-Butanone-d5	3.96	46	316005	48.20	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.40%
24) Chloroform-d	4.40	84	325686	6.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	131.60%#
26) 1,2-Dichloroethane-d4	5.08	65	164886	6.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	139.20%#
32) Benzene-d6	5.09	84	677623	7.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	144.60%#
36) 1,2-Dichloropropane-d6	6.12	67	208176	7.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	140.00%
41) Toluene-d8	7.36	98	542604	6.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	129.20%
43) trans-1,3-Dichloropropene-	7.66	79	60671	5.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	118.60%
46) 2-Hexanone-d5	8.14	63	180708	46.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.32%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	111571	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	175653	7.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	150.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091519\
Data File : VV012808.D
Acq On : 15 Sep 2019 20:35
Operator : SY/MD
Sample : K4865-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMJ8

Quant Time: Sep 16 08:37:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 08:14:10 2019
Response via : Initial Calibration

