

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091519\
 Data File : VV012813.D
 Acq On : 15 Sep 2019 22:48
 Operator : SY/MD
 Sample : K4865-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COML2

Quant Time: Sep 16 09:21:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 08:14:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	389632	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	391971	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	135357	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	198611	7.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	150.40%#
7) Chloroethane-d5	1.58	69	166682	7.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	152.40%#
11) 1,1-Dichloroethene-d2	2.13	63	248382	5.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	116.40%
20) 2-Butanone-d5	3.98	46	410762	64.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	129.16%
24) Chloroform-d	4.40	84	330792	6.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	137.80%#
26) 1,2-Dichloroethane-d4	5.08	65	171021	7.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	148.80%#
32) Benzene-d6	5.09	84	670042	7.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	143.00%#
36) 1,2-Dichloropropane-d6	6.12	67	208389	7.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	140.20%#
41) Toluene-d8	7.36	98	531744	6.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	126.80%
43) trans-1,3-Dichloropropene-	7.66	79	62245	6.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	121.80%
46) 2-Hexanone-d5	8.14	63	222339	57.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.88%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	120150	5.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	176551	7.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	152.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091519\
Data File : VV012813.D
Acq On : 15 Sep 2019 22:48
Operator : SY/MD
Sample : K4865-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COML2

Quant Time: Sep 16 09:21:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 08:14:10 2019
Response via : Initial Calibration

