

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091519\
 Data File : VV012816.D
 Acq On : 16 Sep 2019 00:07
 Operator : SY/MD
 Sample : K4832-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JZ7

Quant Time: Sep 16 09:24:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 08:14:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	397066	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	395594	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	138849	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	199015	7.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	147.80%#
7) Chloroethane-d5	1.58	69	167749	7.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	150.60%#
11) 1,1-Dichloroethene-d2	2.13	63	245540	5.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	113.00%
20) 2-Butanone-d5	3.98	46	407568	62.88	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.76%
24) Chloroform-d	4.40	84	319160	6.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	130.40%#
26) 1,2-Dichloroethane-d4	5.08	65	167749	7.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	143.20%#
32) Benzene-d6	5.09	84	650718	6.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	137.60%#
36) 1,2-Dichloropropane-d6	6.12	67	200935	6.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	134.00%
41) Toluene-d8	7.36	98	513805	6.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.40%
43) trans-1,3-Dichloropropene-	7.66	79	61297	5.94	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	118.80%
46) 2-Hexanone-d5	8.14	63	216001	55.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.58%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	120255	5.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	171543	7.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	144.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091519\
Data File : VV012816.D
Acq On : 16 Sep 2019 00:07
Operator : SY/MD
Sample : K4832-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JZ7

Quant Time: Sep 16 09:24:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 08:14:10 2019
Response via : Initial Calibration

