

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091520\
 Data File : VV018215.D
 Acq On : 15 Sep 2020 11:44
 Operator : SY/MD
 Sample : VV0915WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK69

Quant Time: Sep 16 06:45:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 16 05:45:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	301945	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	295303	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	146411	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	99782	40.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.10%
7) Chloroethane-d5	1.58	69	88128	43.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.54%
11) 1,1-Dichloroethene-d2	2.12	63	150713	33.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.78%
21) 2-Butanone-d5	3.90	46	138708	92.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.09%
24) Chloroform-d	4.37	84	196550	46.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.08%
26) 1,2-Dichloroethane-d4	5.05	65	128262	46.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.74%
32) Benzene-d6	5.07	84	385021	45.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.08%
36) 1,2-Dichloropropane-d6	6.09	67	124573	46.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.76%
41) Toluene-d8	7.33	98	342174	44.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.82%
43) trans-1,3-Dichloropropene-	7.64	79	60907	44.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.24%
47) 2-Hexanone-d5	8.10	63	105293	103.49	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.49%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	169114	50.16	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.32%
64) 1,2-Dichlorobenzene-d4	11.65	152	149864	49.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091520\
Data File : VV018215.D
Acq On : 15 Sep 2020 11:44
Operator : SY/MD
Sample : VV0915WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK69

Quant Time: Sep 16 06:45:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 16 05:45:07 2020
Response via : Initial Calibration

