

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091520\
 Data File : VV018225.D
 Acq On : 15 Sep 2020 16:33
 Operator : SY/MD
 Sample : L4012-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWZ3

Quant Time: Sep 16 07:15:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 16 05:45:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	289046	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	288647	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	142848	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	79460	33.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.46%
7) Chloroethane-d5	1.58	69	77463	40.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.38%
11) 1,1-Dichloroethene-d2	2.12	63	128709	29.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.58%#
21) 2-Butanone-d5	3.90	46	135268	93.82	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.82%
24) Chloroform-d	4.37	84	185750	45.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.90%
26) 1,2-Dichloroethane-d4	5.05	65	125359	47.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.70%
32) Benzene-d6	5.07	84	364751	44.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.28%
36) 1,2-Dichloropropane-d6	6.09	67	120099	45.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.48%
41) Toluene-d8	7.33	98	325605	43.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.48%
43) trans-1,3-Dichloropropene-	7.64	79	56655	42.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.92%
47) 2-Hexanone-d5	8.10	63	100304	100.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.86%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	163034	49.47	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.94%
64) 1,2-Dichlorobenzene-d4	11.65	152	145432	49.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091520\
Data File : VV018225.D
Acq On : 15 Sep 2020 16:33
Operator : SY/MD
Sample : L4012-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWZ3

Quant Time: Sep 16 07:15:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 16 05:45:07 2020
Response via : Initial Calibration

