

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091522\
 Data File : VV027851.D
 Acq On : 15 Sep 2022 13:20
 Operator : SY/MD
 Sample : N4621-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB8G5

Quant Time: Sep 16 01:32:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 16 01:29:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	119006	5.000	ug/L	# 0.03
28) Chlorobenzene-d5	8.850	117	119455	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55648	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	30524	3.891	ug/L	0.01
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.800%	
7) Chloroethane-d5	1.568	69	33162	4.983	ug/L	0.01
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.600%	
11) 1,1-Dichloroethene-d2	2.108	63	55412	3.329	ug/L	0.01
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.600%	
20) 2-Butanone-d5	3.889	46	41156	23.674	ug/L	-0.05
Spiked Amount	50.000	Range 40 - 130	Recovery	=	47.340%	
24) Chloroform-d	4.346	84	85272	5.309	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	106.200%	
26) 1,2-Dichloroethane-d4	5.030	65	41949	4.904	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
32) Benzene-d6	5.047	84	151689	4.377	ug/L	0.02
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.600%	
36) 1,2-Dichloropropane-d6	6.066	67	42918	4.230	ug/L	0.02
Spiked Amount	5.000	Range 60 - 140	Recovery	=	84.600%	
41) Toluene-d8	7.313	98	126338	4.335	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.800%	
43) trans-1,3-Dichloroprop...	7.622	79	13146	3.862	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.200%	
46) 2-Hexanone-d5	8.091	63	46830	37.012	ug/L	-0.02
Spiked Amount	50.000	Range 45 - 130	Recovery	=	74.020%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	28228	4.332	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	86.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	46389	5.075	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.400%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	25674	2.414	ug/L	96
22) cis-1,2-Dichloroethene	3.908	96	87993	9.693	ug/L #	85
34) Trichloroethene	5.911	95	22750	2.304	ug/L	94
51) Chlorobenzene	8.886	112	2723	0.110	ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091522\
Data File : VV027851.D
Acq On : 15 Sep 2022 13:20
Operator : SY/MD
Sample : N4621-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CB8G5

Quant Time: Sep 16 01:32:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 16 01:29:01 2022
Response via : Initial Calibration

